

The comparison of the regulatory limits in this report reflect the current Chemtech Consulting Group Inc. knowledge of the standards and are intended as general guidance for the user. Please consult appropriate regulations and cleanup standards for your specific application.

Sample ID		P001-DW-5027-3	P001-DW-5029-3	P001-DW-6006-3	P001-DW-6009-3	P001-DW-6010-3	P001-DW-6011-3	P001-DW-6017-3	P001-DW-6018-3	P001-DW-6021-3	P001-DW-6024-3
Lab Sample Number		E3848-01	E3848-02	E3848-03	E3848-04	E3848-05	E3848-06	E3848-07	E3848-08	E3848-09	E3848-10
Sampling Date		9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013
Matrix		SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL
Dilution Factor		1	1	1	1	1	1	1	1	1	1
Units		mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg
COMPOUND	CAS #										
Reactive Sulfide	RESU	42	38	35	40	46	42	37	43	42	38

**Qualifiers**

U - The compound was not detected at the indicated concentration.

N (Organics) - Presumptive Evidence of a Compound

N (Inorganics) - The matrix spike recovery was outside control limits

J - Data indicates the presence of a compound that meets the identification criteria. The result is less than the quantitation limit but greater than MDL. The concentration given is an approximate value.

B - The analyte was found in the laboratory blank as well as the sample. This indicates possible laboratory contamination of the environmental sample.

P - For dual column analysis, the percent difference between the quantitated concentrations on the two columns is greater than 40%.

\* (Organics) - For dual column analysis, the lowest quantitated concentration is being reported due to coeluting interference.

\* (Inorganics) - The sample/duplicate %RPD was above the control limit.

E (Organics) - Indicates the analyte's concentration exceeds the calibrated range of the instrument for that specific analysis.

E (Inorganics) - The reported value is estimated because of the presence of interference.

D - The reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.

NR - Not analyzed

The comparison of the regulatory limits in this report reflect the current Chemtech Consulting Group Inc. knowledge of the standards and are intended as general guidance for the user. Please consult appropriate regulations and cleanup standards for your specific application.

Sample ID		P001-DW-5027-3	P001-DW-5029-3	P001-DW-6006-3	P001-DW-6009-3	P001-DW-6010-3	P001-DW-6011-3	P001-DW-6017-3	P001-DW-6018-3	P001-DW-6021-3	P001-DW-6024-3
Lab Sample Number		E3848-01	E3848-02	E3848-03	E3848-04	E3848-05	E3848-06	E3848-07	E3848-08	E3848-09	E3848-10
Sampling Date		9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013
Matrix		SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL
Dilution Factor		1	1	1	1	1	1	1	1	1	1
Units		mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg	mg/Kg
COMPOUND	CAS #										
Reactive Cyanide	RECY	0.05 U	0.05 U	0.05 U	0.05 U	0.05 U	0.05 U	0.05 U	0.05 U	0.05 U	0.05 U

**Qualifiers**  
U - The compound was not detected at the indicated concentration.  
N (Organics) - Presumptive Evidence of a Compound  
N (Inorganics) - The matrix spike recovery was outside control limits  
J - Data indicates the presence of a compound that meets the identification criteria. The result is less than the quantitation limit but greater than MDL.  
The concentration given is an approximate value.  
B - The analyte was found in the laboratory blank as well as the sample. This indicates possible laboratory contamination of the environmental sample.  
P - For dual column analysis, the percent difference between the quantitated concentrations on the two columns is greater than 40%.  
\* (Organics) - For dual column analysis, the lowest quantitated concentration is being reported due to coeluting interference.  
\* (Inorganics) - The sample/duplicate %RPD was above the control limit.  
E (Organics) - Indicates the analyte's concentration exceeds the calibrated range of the instrument for that specific analysis.  
E (Inorganics) - The reported value is estimated because of the presence of interference.  
D - The reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.  
NR - Not analyzed

# CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 789-8922 [www.chemtech.net](http://www.chemtech.net)

The comparison of the regulatory limits in this report reflect the current Chemtech Consulting Group Inc. knowledge of the standards and are intended as general guidance for the user. Please consult appropriate regulations and cleanup standards for your specific application.

<b>Sample ID</b>	<b>P001-DW-6011-3</b>
<b>Lab Sample Number</b>	<b>E3848-06</b>
<b>Sampling Date</b>	<b>9/24/2013</b>
<b>Matrix</b>	<b>SOIL</b>
<b>Dilution Factor</b>	<b>1</b>
<b>Units</b>	<b>o C</b>
<b>COMPOUND</b>	<b>CAS #</b>
<b>Ignitability</b>	<b>IGNIT</b>
	<b>NO</b>

## Qualifiers

U - The compound was not detected at the indicated concentration.

N (Organics) - Presumptive Evidence of a Compound

N (Inorganics) - The matrix spike recovery was outside control limits

J - Data indicates the presence of a compound that meets the identification criteria. The result is less than the quantitation limit but greater than MDL.  
The concentration given is an approximate value.

B - The analyte was found in the laboratory blank as well as the sample. This indicates possible laboratory contamination of the environmental sample.

P - For dual column analysis, the percent difference between the quantitated concentrations on the two columns is greater than 40%.

\* (Organics) - For dual column analysis, the lowest quantitated concentration is being reported due to coeluting interference.

\* (Inorganics) - The sample/duplicate %RPD was above the control limit.

E (Organics) - Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.

E (Inorganics) - The reported value is estimated because of the presence of interference.

D - The reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.

NR - Not analyzed

The comparison of the regulatory limits in this report reflect the current Chemtech Consulting Group Inc. knowledge of the standards and are intended as general guidance for the user. Please consult appropriate regulations and cleanup standards for your specific application.

Sample ID	P001-DW-5027-3	P001-DW-5029-3	P001-DW-6006-3	P001-DW-6009-3	P001-DW-6010-3	P001-DW-6017-3	P001-DW-6018-3	P001-DW-6021-3	P001-DW-6024-3
Lab Sample Number	E3848-01	E3848-02	E3848-03	E3848-04	E3848-05	E3848-07	E3848-08	E3848-09	E3848-10
Sampling Date	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013
Matrix	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL
Dilution Factor									
Units	o F	o F	o F	o F	o F	o F	o F	o F	o F
COMPOUND	CAS #								
Flashpoint	78	84	78	78	76	80	80	78	88

Qualifiers

- U - The compound was not detected at the indicated concentration.
- N (Organics) - Presumptive Evidence of a Compound
- N (Inorganics) - The matrix spike recovery was outside control limits
- J - Data indicates the presence of a compound that meets the identification criteria. The result is less than the quantitation limit but greater than MDL. The concentration given is an approximate value.
- B - The analyte was found in the laboratory blank as well as the sample. This indicates possible laboratory contamination of the environmental sample.
- P - For dual column analysis, the percent difference between the quantitated concentrations on the two columns is greater than 40%.
- \* (Organics) - For dual column analysis, the lowest quantitated concentration is being reported due to coeluting interference.
- \* (Inorganics) - The sample/duplicate %RPD was above the control limit.
- E (Organics) - Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
- E (Inorganics) - The reported value is estimated because of the presence of interference.
- D - The reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
- NR - Not analyzed

The comparison of the regulatory limits in this report reflect the current Chemtech Consulting Group Inc. knowledge of the standards and are intended as general guidance for the user. Please consult appropriate regulations and cleanup standards for your specific application.

Sample ID		P001-DW-5027-3	P001-DW-5029-3	P001-DW-6006-3	P001-DW-6009-3	P001-DW-6010-3	P001-DW-6011-3	P001-DW-6017-3	P001-DW-6018-3	P001-DW-6021-3	P001-DW-6024-3
Lab Sample Number		E3848-01	E3848-02	E3848-03	E3848-04	E3848-05	E3848-06	E3848-07	E3848-08	E3848-09	E3848-10
Sampling Date		9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013	9/24/2013
Matrix		SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL	SOIL
Dilution Factor		1	1	1	1	1	1	1	1	1	1
Units		pH	pH	pH	pH	pH	pH	pH	pH	pH	pH
COMPOUND	CAS #										
Corrosivity (as pH)	CORR	3.29	3.62	4	6.51	6.39	4.38	10.2	12.3	13.1	8.07

**Qualifiers**  
U - The compound was not detected at the indicated concentration.  
N (Organics) - Presumptive Evidence of a Compound  
N (Inorganics) - The matrix spike recovery was outside control limits  
J - Data indicates the presence of a compound that meets the identification criteria. The result is less than the quantitation limit but greater than MDL.  
The concentration given is an approximate value.  
B - The analyte was found in the laboratory blank as well as the sample. This indicates possible laboratory contamination of the environmental sample.  
P - For dual column analysis, the percent difference between the quantitated concentrations on the two columns is greater than 40%.  
\* (Organics) - For dual column analysis, the lowest quantitated concentration is being reported due to coeluting interference.  
\* (Inorganics) - The sample/duplicate %RPD was above the control limit.  
E (Organics) - Indicates the analyte's concentration exceeds the calibrated range of the instrument for that specific analysis.  
E (Inorganics) - The reported value is estimated because of the presence of interference.  
D - The reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.  
NR - Not analyzed